



1. What are the clinical requirements for successful metal-free restorations?
  - a. Resistance, adaptation and color stability
  - b. Surface texture, anatomy and adaptation
  - c. Volume; glaze and surface texture
  - d. Anatomy, color stability and glaze
2. The Procera system is based on CAD/CAM technology for industrial production of porcelain crown infrastructure using alumina, zirconia and titanium.
  - a. True
  - b. False
3. Study models are not important for clinical planning.
  - a. True
  - b. False
4. In which countries are there Procera production centers?
  - a. England and Germany
  - b. Brazil and Russia
  - c. Netherlands and Brazil
  - d. Sweden and the United States
5. The clinical exam and the anamnesis are very important to determine a diagnosis and a treatment plan.
  - a. True
  - b. False
6. The manufacture of provision restorations is very important.
  - a. True
  - b. False
7. The selection of the ceramic for the coping is based on TEC (thermal expansion coefficient), which must be close to the aluminum oxide CET, i.e.,  $7.3 \times 10^{-6} \text{ }^{\circ}\text{C}$  necessary.
  - a. True
  - b. False
8. One of the most important steps for long-term restorations is providing information to the patient on how to clean the restorations.
  - a. True
  - b. False
9. The aluminum oxide frameworks may be cemented with zinc phosphate cement.
  - a. True
  - b. False
10. The aluminum oxide coping with microfractures and the non-adaptation of the finishing line of the preparation contribute for the success of the proposed treatment, as well as for the restoration longevity.
  - a. True
  - b. False

Passing quiz grades are worth ½ point documented scientific credit. Quiz results will appear on the NBC CDT Online Education Tracking System at [www.nadl.org](http://www.nadl.org) under the Certification Section, which is updated on or around the 15th of every month. Once completed, send your quiz to the National Board for Certification in Dental Laboratory Technology at the address or fax number below. Name: \_\_\_\_\_ CDT #: \_\_\_\_\_ Date: \_\_\_\_\_ 325 John Knox Rd, Suite L103, Tallahassee, Florida 32303 (850) 222-0053 FAX This quiz is provided to test the technician's comprehension of the article's content, and does not necessarily serve as an endorsement of the content by NADL or NBC.